

## **Press release: Alert to users of FreeStyle Libre flash glucose monitoring system regarding skin reactions to sensor adhesive**

Use of barrier methods to reduce skin reactions may affect device performance.