

[Press release: Diabetes test strips recalled in new alert](#)

People with diabetes are advised to stop using and return specific lots of Accu-Chek Aviva and Accu-Chek Performa test strips following a recent recall by the manufacturer.

Accu-Chek Inform II test strips have also been recalled but are supplied in the UK by Roche for professional use only.

The test strips, commonly used by diabetics for blood glucose testing, may give increased strip error messages prior to dosing with blood and in some cases may give falsely high or low readings which may be hard to detect.

The Medicines and Healthcare products Regulatory Agency (MHRA) are urging users to check the lot numbers of their test strips against the lot numbers listed in the table below. More information can be found in this [field safety notice](#). It is estimated that more than 260,000 packs have been affected.

If anyone finds they have test strips from the affected lots, they are advised to seek alternative testing methods and return affected lots to their pharmacy or shop where they will be offered a replacement.

It is also advised anyone with concerns about their blood glucose readings should discuss this with a healthcare professional.

Affected products

ACCU-CHEK AVIVA STRIPS (50s)

- 497392
- 497391
- 496915
- 496809
- 496802
- 496807

ACCU-CHEK AVIVA STRIPS (10s)

- 497344
- 497392

ACCU-CHEK PERFORMA STRIPS (10s)

- 476597
- 476646

Accu-Chek Inform II strips (professional use only)

- 476614

Accu-Chek Performa Nano mmol Kit

(Please note the FSN only applies to the pack of Performa 10 test strips contained within the kit)

- 10153116
- 10153114
- 10153115
- 10153112
- 10153111

John Wilkinson, MHRA's Director of Medical Devices said:

It is important people check their test strips and if necessary seek alternatives as soon as possible.

If people have any questions about their blood glucose readings when using these test strips and meters they should speak with their doctor or pharmacist.

We continue to encourage people to report any issues involving medical devices to MHRA via our [Yellow Card Scheme](#).

Dan Howarth, Head of Care at Diabetes UK said:

It is incredibly important that people living with diabetes are able to rely on the technology that is designed to help them manage their blood sugar, so we'd strongly recommend that anyone using these test strips check their batch numbers and get replacement strips accordingly.

Inaccurate readings could mean you over- or under-dose with insulin, which in the short term can cause problems with hyper and hypoglycaemia. If you're worried about your readings, we'd recommend speaking to your GP or a pharmacist as soon as possible.

Media enquiries

News centre
MHRA

151 Buckingham Palace Road

Victoria
London
SW1W 9SZ

Email

newscentre@mhra.gov.uk

During office hours:

020 3080 7651 (08:30 – 17:00)

Out of office hours:

07770 446 189 (17:00 – 08:30)

Office hours are Monday to Friday, 8:30am to 5pm. For real-time updates including the latest press releases and news statements, see our Twitter channel at <https://www.twitter.com/mhrapress>